

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073119\
 Data File : VN057005.D
 Acq On : 31 Jul 2019 11:34
 Operator : JC/SP
 Sample : PB121844ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121844ZHE#02

Quant Time: Aug 01 01:44:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	567331	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	825093	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	714164	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	204908	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	283616	48.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.68%	
35) Dibromofluoromethane	7.58	113	258677	48.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.94%	
50) Toluene-d8	10.08	98	984992	48.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.80%	
62) 4-Bromofluorobenzene	12.40	95	267069	40.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.46%	
Target Compounds						
20) Methylene Chloride	4.54	84	10258	1.693	ug/l	88
43) Isopropyl Acetate	8.16	43	22012	2.290	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073119\
 Data File : VN057005.D
 Acq On : 31 Jul 2019 11:34
 Operator : JC/SP
 Sample : PB121844ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121844ZHE#02

Quant Time: Aug 01 01:44:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

