

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073119\
 Data File : VN057010.D
 Acq On : 31 Jul 2019 13:31
 Operator : JC/SP
 Sample : PB121844ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB121844ZHE#06

Quant Time: Aug 01 01:55:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	562408	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	830126	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	713875	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	213408	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	281992	48.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.96%	
35) Dibromofluoromethane	7.58	113	251618	47.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.70%	
50) Toluene-d8	10.08	98	997818	49.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.48%	
62) 4-Bromofluorobenzene	12.40	95	273117	41.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.80%	
Target Compounds						
20) Methylene Chloride	4.55	84	23414	3.897	ug/l	97
43) Isopropyl Acetate	8.15	43	19894	2.057	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073119\
Data File : VN057010.D
Acq On : 31 Jul 2019 13:31
Operator : JC/SP
Sample : PB121844ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB121844ZHE#06

Quant Time: Aug 01 01:55:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

