

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073119\
 Data File : VN057014.D
 Acq On : 31 Jul 2019 15:04
 Operator : JC/SP
 Sample : K4077-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20190612-LOC-13

Quant Time: Aug 01 02:02:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	556069	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	812749	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	682042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	206789	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	273207	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
35) Dibromofluoromethane	7.58	113	247051	47.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.96%	
50) Toluene-d8	10.08	98	959411	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
62) 4-Bromofluorobenzene	12.40	95	262229	40.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.20%	
Target Compounds						
16) Acetone	3.81	43	25689	12.457	ug/l	93
20) Methylene Chloride	4.54	84	137787	23.196	ug/l	97
43) Isopropyl Acetate	8.17	43	26164	2.764	ug/l #	86
95) Naphthalene	15.12	128	6692	4.295	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073119\
Data File : VN057014.D
Acq On : 31 Jul 2019 15:04
Operator : JC/SP
Sample : K4077-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20190612-LOC-13

Quant Time: Aug 01 02:02:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

