

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073119\
 Data File : VN057015.D
 Acq On : 31 Jul 2019 15:27
 Operator : JC/SP
 Sample : K4066-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-13

Quant Time: Aug 01 02:04:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	539876	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	801538	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	675710	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	194054	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	267481	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
35) Dibromofluoromethane	7.58	113	243609	47.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.96%	
50) Toluene-d8	10.08	98	946568	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
62) 4-Bromofluorobenzene	12.40	95	255716	40.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.28%	
Target Compounds						
16) Acetone	3.81	43	27878	13.924	ug/l	95
18) Methyl Acetate	4.32	43	8067	1.263	ug/l #	87
20) Methylene Chloride	4.54	84	24294	4.213	ug/l	93
43) Isopropyl Acetate	8.16	43	28903	3.096	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073119\
Data File : VN057015.D
Acq On : 31 Jul 2019 15:27
Operator : JC/SP
Sample : K4066-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampled :
TP-13

Quant Time: Aug 01 02:04:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

