

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073120\
 Data File : VN062679.D
 Acq On : 31 Jul 2020 12:40
 Operator : JC/MD
 Sample : PB130640ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130640ZHE#02

Quant Time: Aug 01 06:39:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	134762	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	240291	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	249234	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	89182	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	116600	55.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.40%	
35) Dibromofluoromethane	7.55	113	81358	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
50) Toluene-d8	10.06	98	325421	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.34%	
62) 4-Bromofluorobenzene	12.38	95	117481	51.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.04%	
Target Compounds						
8) Diethyl Ether	3.37	74	1486	1.432	ug/l	94
16) Acetone	3.78	43	14742	18.713	ug/l	99
18) Methyl Acetate	4.28	43	4231	2.122	ug/l #	79
20) Methylene Chloride	4.50	84	15813	8.369	ug/l	89
43) Isopropyl Acetate	8.13	43	48049	10.796	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073120\
Data File : VN062679.D
Acq On : 31 Jul 2020 12:40
Operator : JC/MD
Sample : PB130640ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130640ZHE#02

Quant Time: Aug 01 06:39:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

