

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073120\
 Data File : VN062687.D
 Acq On : 31 Jul 2020 15:54
 Operator : JC/MD
 Sample : PB130640ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130640ZHE#10

Quant Time: Aug 01 06:58:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	126482	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	240150	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	235599	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	77540	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	111970	56.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.98%	
35) Dibromofluoromethane	7.55	113	79876	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
50) Toluene-d8	10.06	98	311851	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
62) 4-Bromofluorobenzene	12.38	95	111076	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
Target Compounds						
16) Acetone	3.78	43	9815	13.275	ug/l	97
18) Methyl Acetate	4.29	43	7234	3.865	ug/l	92
20) Methylene Chloride	4.51	84	6769	3.817	ug/l #	93
43) Isopropyl Acetate	8.13	43	50134	11.271	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073120\
Data File : VN062687.D
Acq On : 31 Jul 2020 15:54
Operator : JC/MD
Sample : PB130640ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130640ZHE#10

Quant Time: Aug 01 06:58:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

