

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073120\
 Data File : VN062688.D
 Acq On : 31 Jul 2020 16:18
 Operator : JC/MD
 Sample : PB130640ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130640ZHE#11

Quant Time: Aug 01 06:59:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	127868	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	245058	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	241529	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	86706	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	112898	56.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.68%	
35) Dibromofluoromethane	7.55	113	79762	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
50) Toluene-d8	10.06	98	315484	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
62) 4-Bromofluorobenzene	12.38	95	116084	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
Target Compounds						
16) Acetone	3.79	43	9787	13.093	ug/l	90
18) Methyl Acetate	4.29	43	7703	4.071	ug/l #	72
20) Methylene Chloride	4.50	84	7502	4.185	ug/l	98
43) Isopropyl Acetate	8.13	43	50306	11.083	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073120\
Data File : VN062688.D
Acq On : 31 Jul 2020 16:18
Operator : JC/MD
Sample : PB130640ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130640ZHE#11

Quant Time: Aug 01 06:59:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

