

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073120\
 Data File : VN062704.D
 Acq On : 31 Jul 2020 23:33
 Operator : JC/MD
 Sample : L3518-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Quant Time: Aug 03 16:27:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	117706	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	210876	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	197959	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	72358	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	100402	54.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.82%	
35) Dibromofluoromethane	7.55	113	69663	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
50) Toluene-d8	10.06	98	268796	48.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.28%	
62) 4-Bromofluorobenzene	12.38	95	95616	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
Target Compounds						
16) Acetone	3.78	43	10097	14.674	ug/l	93
20) Methylene Chloride	4.51	84	10869	6.586	ug/l	88
95) Naphthalene	15.11	128	4453	5.041	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073120\
Data File : VN062704.D
Acq On : 31 Jul 2020 23:33
Operator : JC/MD
Sample : L3518-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Quant Time: Aug 03 16:27:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

