

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073120\
 Data File : VN062705.D
 Acq On : 31 Jul 2020 23:58
 Operator : JC/MD
 Sample : L3518-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-6

Quant Time: Aug 01 06:30:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	141995	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	252573	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	250228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	88876	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	120657	54.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.40%	
35) Dibromofluoromethane	7.55	113	85677	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
50) Toluene-d8	10.06	98	332775	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
62) 4-Bromofluorobenzene	12.38	95	117916	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
Target Compounds						
16) Acetone	3.78	43	19530	23.528	ug/l	95
20) Methylene Chloride	4.50	84	11191	5.621	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073120\
Data File : VN062705.D
Acq On : 31 Jul 2020 23:58
Operator : JC/MD
Sample : L3518-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-6

Quant Time: Aug 01 06:30:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

