

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN073120\
 Data File : VN062714.D
 Acq On : 1 Aug 2020 3:35
 Operator : JC/MD
 Sample : L3518-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Manual Integrations
 APPROVED

MMDadoda
 8/3/2020 5:03:21 PM

Quant Time: Aug 01 06:47:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	145486	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	260163	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	274709	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	100862	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	125964	55.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.46%	
35) Dibromofluoromethane	7.55	113	88643	51.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.74%	
50) Toluene-d8	10.06	98	350567	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
62) 4-Bromofluorobenzene	12.38	95	131615	52.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.58%	
Target Compounds						
16) Acetone	3.79	43	9285	10.917	ug/l	98
18) Methyl Acetate	4.28	43	4251m	1.974	ug/l	
20) Methylene Chloride	4.50	84	7781	3.815	ug/l	93
43) Isopropyl Acetate	8.13	43	47014	9.756	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN073120\
Data File : VN062714.D
Acq On : 1 Aug 2020 3:35
Operator : JC/MD
Sample : L3518-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Manual Integrations
APPROVED

MMDadoda
8/3/2020 5:03:21 PM

Quant Time: Aug 01 06:47:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

