

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080118\
 Data File : VN050240.D
 Acq On : 1 Aug 2018 11:25
 Operator : MD\SY
 Sample : J4256-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17

Manual Integrations
 APPROVED

MMDadoda
 8/2/2018 9:53:50 AM

Quant Time: Aug 04 00:05:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	695203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1130706	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	949306	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	390084	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	485768	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
35) Dibromofluoromethane	7.59	113	428312	46.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.78%	
50) Toluene-d8	10.09	98	1536533	45.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.54%	
62) 4-Bromofluorobenzene	12.40	95	452137	40.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.92%	
Target Compounds						
16) Acetone	3.83	43	10501	0.258	ug/l	Qvalue 100
19) Methyl tert-butyl Ether	5.05	73	8198m	0.384	ug/l	
73) Isopropylbenzene	12.25	105	33369	1.103	ug/l	99
78) n-propylbenzene	12.59	91	53094	1.570	ug/l	98
85) sec-Butylbenzene	13.17	105	29877	1.077	ug/l	89
89) n-Butylbenzene	13.62	91	8006m	0.416	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080118\
Data File : VN050240.D
Acq On : 1 Aug 2018 11:25
Operator : MD\SY
Sample : J4256-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-17

Manual Integrations
APPROVED

MMDadoda
8/2/2018 9:53:50 AM

Quant Time: Aug 04 00:05:03 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

