

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080118\
 Data File : VN050247.D
 Acq On : 1 Aug 2018 14:20
 Operator : MD\SY
 Sample : PB111649TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111649TB

Quant Time: Aug 02 02:08:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	693963	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1135488	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	963210	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	372502	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	485932	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
35) Dibromofluoromethane	7.59	113	430822	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
50) Toluene-d8	10.09	98	1541905	45.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.48%	
62) 4-Bromofluorobenzene	12.40	95	437929	39.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.04%	
Target Compounds						
16) Acetone	3.83	43	27316	7.770	ug/l	97
18) Methyl Acetate	4.33	43	26087	1.355	ug/l	93
43) Isopropyl Acetate	8.17	43	496857	29.084	ug/l #	87
59) 2-Hexanone	10.77	43	24381	8.880	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080118\
Data File : VN050247.D
Acq On : 1 Aug 2018 14:20
Operator : MD\SY
Sample : PB111649TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111649TB

Quant Time: Aug 02 02:08:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

