

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080118\
 Data File : VN050248.D
 Acq On : 1 Aug 2018 14:45
 Operator : MD\SY
 Sample : PB111649ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111649ZHE#01

Quant Time: Aug 02 02:11:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	642721	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1071401	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	901516	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	336670	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	461093	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
35) Dibromofluoromethane	7.59	113	405831	46.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.78%	
50) Toluene-d8	10.09	98	1438827	44.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.48%	
62) 4-Bromofluorobenzene	12.40	95	399508	37.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.46%	
Target Compounds						
16) Acetone	3.83	43	27422	8.793	ug/l	94
18) Methyl Acetate	4.34	43	25510	1.525	ug/l	98
43) Isopropyl Acetate	8.17	43	517207	32.207	ug/l #	86
59) 2-Hexanone	10.77	43	24247	9.080	ug/l #	56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080118\
Data File : VN050248.D
Acq On : 1 Aug 2018 14:45
Operator : MD\SY
Sample : PB111649ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111649ZHE#01

Quant Time: Aug 02 02:11:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

