

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080119\
 Data File : VN057026.D
 Acq On : 1 Aug 2019 10:36
 Operator : JC/SP
 Sample : K4079-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TANK-4

Quant Time: Aug 02 05:27:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N072419W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 24 04:19:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.18	128	76022	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	381638	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.40	117	290922	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.02	65	145476	29.24	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	97.47%
60) 4-Bromofluorobenzene	12.40	95	89265	20.47	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	68.23%
63) Toluene-d8	10.08	98	452200	32.91	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	109.70%

Target Compounds					Ovalue
15) Acetone	3.81	58	8390	16.941	ug/l 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080119\
Data File : VN057026.D
Acq On : 1 Aug 2019 10:36
Operator : JC/SP
Sample : K4079-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TANK-4

Quant Time: Aug 02 05:27:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N072419W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jul 24 04:19:57 2019
Response via : Initial Calibration

