

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080119\
 Data File : VN057031.D
 Acq On : 1 Aug 2019 12:52
 Operator : JC/SP
 Sample : K4097-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB203-EB-20190730

Quant Time: Aug 02 05:37:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	574290	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	835503	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	677001	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	174318	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	323841	55.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.18%	
35) Dibromofluoromethane	7.58	113	298527	55.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.62%	
50) Toluene-d8	10.08	98	1108415	54.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.68%	
62) 4-Bromofluorobenzene	12.40	95	275982	41.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.12%	
Target Compounds						
16) Acetone	3.81	43	14618	6.864	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080119\
Data File : VN057031.D
Acq On : 1 Aug 2019 12:52
Operator : JC/SP
Sample : K4097-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB203-EB-20190730

Quant Time: Aug 02 05:37:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

