

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080119\
 Data File : VN057033.D
 Acq On : 1 Aug 2019 13:38
 Operator : JC/SP
 Sample : K4097-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB203-GW-443-445

Quant Time: Aug 02 11:26:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	570930	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	832694	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	678711	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	180939	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	327101	55.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.94%	
35) Dibromofluoromethane	7.58	113	298911	56.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.14%	
50) Toluene-d8	10.08	98	1128567	55.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.04%	
62) 4-Bromofluorobenzene	12.40	95	289254	43.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.42%	
Target Compounds						
16) Acetone	3.81	43	11345	5.358	ug/l	95
27) cis-1,2-Dichloroethene	6.81	96	6479	1.028	ug/l	92
44) Trichloroethene	8.83	130	3247	0.522	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080119\
Data File : VN057033.D
Acq On : 1 Aug 2019 13:38
Operator : JC/SP
Sample : K4097-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB203-GW-443-445

Quant Time: Aug 02 11:26:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

