

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080119\
 Data File : VN057041.D
 Acq On : 1 Aug 2019 16:44
 Operator : JC/SP
 Sample : K4097-09RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB203-EB-20190730RE

Quant Time: Aug 02 06:11:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	560137	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	812617	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	666007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	171210	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	313864	54.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.48%	
35) Dibromofluoromethane	7.58	113	286381	55.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.10%	
50) Toluene-d8	10.08	98	1075506	54.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.42%	
62) 4-Bromofluorobenzene	12.40	95	268738	41.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.22%	
Target Compounds						
16) Acetone	3.82	43	13884	6.684	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080119\
Data File : VN057041.D
Acq On : 1 Aug 2019 16:44
Operator : JC/SP
Sample : K4097-09RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB203-EB-20190730RE

Quant Time: Aug 02 06:11:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

