

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
 Data File : VN073653.D
 Acq On : 01 Aug 2022 12:06
 Operator : JC\MD
 Sample : VN0801WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0801WBL01

Quant Time: Aug 02 04:58:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	374725	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	601634	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	567163	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	266369	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	218771	47.906	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.820%
35) Dibromofluoromethane	7.951	113	207017	54.797	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.600%
50) Toluene-d8	10.381	98	667893	44.926	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.860%
62) 4-Bromofluorobenzene	12.674	95	242939	49.086	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
Data File : VN073653.D
Acq On : 01 Aug 2022 12:06
Operator : JC\MD
Sample : VN0801WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0801WBL01

Quant Time: Aug 02 04:58:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

