

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
 Data File : VN073656.D
 Acq On : 01 Aug 2022 13:41
 Operator : JC\MD
 Sample : N3977-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MS-TOTE-0729

Quant Time: Aug 02 05:00:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	370440	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	578816	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	538261	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	247728	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	214193	47.447	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.900%
35) Dibromofluoromethane	7.951	113	199200	54.807	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.620%
50) Toluene-d8	10.381	98	656483	45.900	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.800%
62) 4-Bromofluorobenzene	12.674	95	233912	49.119	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.240%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.287	59	16721	17.174	ug/l	96
16) Acetone	4.228	43	136568	56.323	ug/l	98
95) Naphthalene	15.439	128	16157	1.148	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
 Data File : VN073656.D
 Acq On : 01 Aug 2022 13:41
 Operator : JC\MD
 Sample : N3977-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MS-TOTE-0729

Quant Time: Aug 02 05:00:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

