

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
 Data File : VN073660.D
 Acq On : 01 Aug 2022 15:19
 Operator : JC\MD
 Sample : N3974-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Aug 02 05:01:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	359070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	580209	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	558697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	270621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	217275	49.653	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.300%
35) Dibromofluoromethane	7.951	113	200370	54.996	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.000%
50) Toluene-d8	10.380	98	666109	46.461	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.920%
62) 4-Bromofluorobenzene	12.674	95	249417	51.725	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.440%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.775	74	3235	1.372	ug/l	76
16) Acetone	4.240	43	13550	5.765	ug/l	92
20) Methylene Chloride	5.016	84	35283	7.402	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
Data File : VN073660.D
Acq On : 01 Aug 2022 15:19
Operator : JC\MD
Sample : N3974-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Aug 02 05:01:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

