

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
 Data File : VN073669.D
 Acq On : 01 Aug 2022 18:59
 Operator : JC\MD
 Sample : N3981-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAT-TOTES-0729

Quant Time: Aug 02 05:05:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	346123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	561651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	532236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	243858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	207729	49.247	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.500%
35) Dibromofluoromethane	7.951	113	193763	54.940	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.880%
50) Toluene-d8	10.381	98	645707	46.526	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.060%
62) 4-Bromofluorobenzene	12.674	95	232020	50.033	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.060%
Target Compounds						
16) Acetone	4.246	43	6003	2.650	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
Data File : VN073669.D
Acq On : 01 Aug 2022 18:59
Operator : JC\MD
Sample : N3981-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAT-TOTES-0729

Quant Time: Aug 02 05:05:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

