

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
 Data File : VN073652.D
 Acq On : 01 Aug 2022 11:26
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 02 04:57:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.485	0.440	9.3	89	0.00
3 P	Chloromethane	0.699	0.609	12.9	91	0.00
4 C	Vinyl Chloride	0.848	0.765	9.8#	89	0.00
5 T	Bromomethane	0.492	0.474	3.7	95	0.00
6 T	Chloroethane	0.592	0.538	9.1	96	0.00
7 T	Trichlorofluoromethane	0.874	0.804	8.0	87	0.00
8 T	Diethyl Ether	0.328	0.307	6.4	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.484	0.484	0.0	100	0.00
10 T	Methyl Iodide	0.492	0.562	-14.2	97	0.00
11 T	Tert butyl alcohol	0.131	0.115	12.2	80	-0.01
12 CM	1,1-Dichloroethene	0.474	0.453	4.4#	95	0.00
13 T	Acrolein	0.130	0.150	-15.4	101	0.00
14 T	Allyl chloride	0.758	0.702	7.4	89	0.00
15 T	Acrylonitrile	0.364	0.353	3.0	91	0.00
16 T	Acetone	0.327	0.282	13.8	88	0.00
17 T	Carbon Disulfide	1.258	1.182	6.0	92	0.00
18 T	Methyl Acetate	0.814	0.758	6.9	92	-0.01
19 T	Methyl tert-butyl Ether	1.621	1.669	-3.0	95	0.00
20 T	Methylene Chloride	0.691	0.569	17.7	96	0.00
21 T	trans-1,2-Dichloroethene	0.514	0.518	-0.8	97	0.00
22 T	Diisopropyl ether	1.636	1.641	-0.3	94	0.00
23 T	Vinyl Acetate	1.372	1.430	-4.2	93	0.00
24 P	1,1-Dichloroethane	1.007	0.962	4.5	96	0.00
25 T	2-Butanone	0.508	0.486	4.3	91	0.00
26 T	2,2-Dichloropropane	0.774	0.750	3.1	95	0.00
27 T	cis-1,2-Dichloroethene	0.633	0.636	-0.5	97	0.00
28 T	Bromochloromethane	0.423	0.440	-4.0	98	0.00
29 T	Tetrahydrofuran	0.321	0.315	1.9	89	0.00
30 C	Chloroform	1.049	1.043	0.6#	98	0.00
31 T	Cyclohexane	0.907	0.854	5.8	94	0.00
32 T	1,1,1-Trichloroethane	0.871	0.859	1.4	95	0.00
33 S	1,2-Dichloroethane-d4	0.609	0.604	0.8	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.314	0.333	-6.1	98	0.00
36 T	1,1-Dichloropropene	0.447	0.488	-9.2	98	0.00
37 T	Ethyl Acetate	0.589	0.597	-1.4	91	0.00
38 T	Carbon Tetrachloride	0.454	0.490	-7.9	100	0.00
39 T	Methylcyclohexane	0.535	0.607	-13.5	98	0.00
40 TM	Benzene	1.453	1.527	-5.1	96	0.00
41 T	Methacrylonitrile	0.267	0.265	0.7	85	0.00
42 TM	1,2-Dichloroethane	0.490	0.501	-2.2	96	0.00
43 T	Isopropyl Acetate	0.796	0.841	-5.7	91	0.00
44 TM	Trichloroethene	0.368	0.390	-6.0	97	0.00
45 C	1,2-Dichloropropane	0.363	0.379	-4.4#	96	0.00
46 T	Dibromomethane	0.252	0.267	-6.0	96	0.00
47 T	Bromodichloromethane	0.484	0.519	-7.2	98	0.00
48 T	Methyl methacrylate	0.348	0.377	-8.3	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
 Data File : VN073652.D
 Acq On : 01 Aug 2022 11:26
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 02 04:57:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	85	0.00
50 S	Toluene-d8	1.236	1.358	-9.9	99	0.00
51 T	4-Methyl-2-Pentanone	0.571	0.615	-7.7	92	0.00
52 CM	Toluene	0.888	0.990	-11.5#	96	0.00
53 T	t-1,3-Dichloropropene	0.478	0.543	-13.6	97	0.00
54 T	cis-1,3-Dichloropropene	0.539	0.595	-10.4	95	0.00
55 T	1,1,2-Trichloroethane	0.370	0.405	-9.5	98	0.00
56 T	Ethyl methacrylate	0.510	0.618	-21.2	95	0.00
57 T	1,3-Dichloropropane	0.628	0.680	-8.3	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.229	0.274	-19.7	95	0.00
59 T	2-Hexanone	0.433	0.474	-9.5	90	0.00
60 T	Dibromochloromethane	0.368	0.426	-15.8	99	0.00
61 T	1,2-Dibromoethane	0.381	0.421	-10.5	97	0.00
62 S	4-Bromofluorobenzene	0.382	0.449	-17.5	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.397	0.418	-5.3	100	0.00
65 PM	Chlorobenzene	1.062	1.130	-6.4	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.411	-7.9	97	0.00
67 C	Ethyl Benzene	1.792	2.007	-12.0#	97	0.00
68 T	m/p-Xylenes	0.683	0.791	-15.8	99	0.00
69 T	o-Xylene	0.666	0.783	-17.6	98	0.00
70 T	Styrene	1.066	1.307	-22.6	100	0.00
71 P	Bromoform	0.314	0.361	-15.0	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.660	3.944	-7.8	99	0.00
74 T	N-amyl acetate	1.383	1.443	-4.3	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.412	1.316	6.8	96	0.00
76 T	1,2,3-Trichloropropane	1.135	0.950	16.3	92	0.00
77 T	Bromobenzene	0.953	0.984	-3.3	99	0.00
78 T	n-propylbenzene	4.118	4.616	-12.1	100	0.00
79 T	2-Chlorotoluene	2.570	2.689	-4.6	99	0.00
80 T	1,3,5-Trimethylbenzene	3.012	3.321	-10.3	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.386	0.403	-4.4	96	0.00
82 T	4-Chlorotoluene	2.436	2.576	-5.7	98	0.00
83 T	tert-Butylbenzene	2.668	2.893	-8.4	96	0.00
84 T	1,2,4-Trimethylbenzene	2.962	3.327	-12.3	99	0.00
85 T	sec-Butylbenzene	3.687	4.159	-12.8	99	0.00
86 T	p-Isopropyltoluene	2.907	3.514	-20.9	101	0.00
87 T	1,3-Dichlorobenzene	1.740	1.811	-4.1	100	0.00
88 T	1,4-Dichlorobenzene	1.741	1.797	-3.2	100	0.00
89 T	n-Butylbenzene	2.365	2.792	-18.1	101	0.00
90 T	Hexachloroethane	0.577	0.601	-4.2	98	0.00
91 T	1,2-Dichlorobenzene	1.763	1.824	-3.5	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.284	0.283	0.4	95	0.00
93 T	1,2,4-Trichlorobenzene	0.876	1.021	-16.6	100	0.00
94 T	Hexachlorobutadiene	0.476	0.507	-6.5	100	0.00
95 T	Naphthalene	2.841	3.317	-16.8	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
Data File : VN073652.D
Acq On : 01 Aug 2022 11:26
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 02 04:57:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.921	1.040	-12.9	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6