

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
 Data File : VN073674.D
 Acq On : 01 Aug 2022 21:02
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 02 05:07:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	0.485	0.440	9.3	103	0.00
3 P	Chloromethane	0.699	0.550	21.3	95	0.00
4 C	Vinyl Chloride	0.848	0.716	15.6#	96	0.00
5 T	Bromomethane	0.492	0.414	15.9	96	0.00
6 T	Chloroethane	0.592	0.469	20.8	97	0.00
7 T	Trichlorofluoromethane	0.874	0.747	14.5	93	0.00
8 T	Diethyl Ether	0.328	0.308	6.1	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.484	0.438	9.5	105	0.00
10 T	Methyl Iodide	0.492	0.556	-13.0	111	0.00
11 T	Tert butyl alcohol	0.131	0.123	6.1	100	0.00
12 CM	1,1-Dichloroethene	0.474	0.431	9.1#	105	0.00
13 T	Acrolein	0.130	0.134	-3.1	104	0.00
14 T	Allyl chloride	0.758	0.671	11.5	98	0.00
15 T	Acrylonitrile	0.364	0.336	7.7	100	0.00
16 T	Acetone	0.327	0.277	15.3	101	0.00
17 T	Carbon Disulfide	1.258	1.071	14.9	97	0.00
18 T	Methyl Acetate	0.814	0.749	8.0	105	0.00
19 T	Methyl tert-butyl Ether	1.621	1.596	1.5	106	0.00
20 T	Methylene Chloride	0.691	0.531	23.2	104	0.00
21 T	trans-1,2-Dichloroethene	0.514	0.479	6.8	104	0.00
22 T	Diisopropyl ether	1.636	1.538	6.0	102	0.00
23 T	Vinyl Acetate	1.372	1.343	2.1	101	0.00
24 P	1,1-Dichloroethane	1.007	0.895	11.1	103	0.00
25 T	2-Butanone	0.508	0.460	9.4	99	0.00
26 T	2,2-Dichloropropane	0.774	0.614	20.7	91	0.00
27 T	cis-1,2-Dichloroethene	0.633	0.584	7.7	103	0.00
28 T	Bromochloromethane	0.423	0.353	16.5	91	0.00
29 T	Tetrahydrofuran	0.321	0.302	5.9	98	0.00
30 C	Chloroform	1.049	0.942	10.2#	102	0.00
31 T	Cyclohexane	0.907	0.777	14.3	99	0.00
32 T	1,1,1-Trichloroethane	0.871	0.810	7.0	103	0.00
33 S	1,2-Dichloroethane-d4	0.609	0.586	3.8	109	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	107	0.00
35 S	Dibromofluoromethane	0.314	0.316	-0.6	112	0.00
36 T	1,1-Dichloropropene	0.447	0.424	5.1	102	0.00
37 T	Ethyl Acetate	0.589	0.548	7.0	100	0.00
38 T	Carbon Tetrachloride	0.454	0.428	5.7	104	0.00
39 T	Methylcyclohexane	0.535	0.523	2.2	101	0.00
40 TM	Benzene	1.453	1.349	7.2	102	0.00
41 T	Methacrylonitrile	0.267	0.267	0.0	102	0.00
42 TM	1,2-Dichloroethane	0.490	0.446	9.0	102	0.00
43 T	Isopropyl Acetate	0.796	0.777	2.4	101	0.00
44 TM	Trichloroethene	0.368	0.343	6.8	102	0.00
45 C	1,2-Dichloropropane	0.363	0.334	8.0#	101	0.00
46 T	Dibromomethane	0.252	0.237	6.0	102	0.00
47 T	Bromodichloromethane	0.484	0.453	6.4	103	0.00
48 T	Methyl methacrylate	0.348	0.345	0.9	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
 Data File : VN073674.D
 Acq On : 01 Aug 2022 21:02
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 02 05:07:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	100	0.00
50 S	Toluene-d8	1.236	1.255	-1.5	109	0.00
51 T	4-Methyl-2-Pentanone	0.571	0.554	3.0	99	0.00
52 CM	Toluene	0.888	0.876	1.4#	102	0.00
53 T	t-1,3-Dichloropropene	0.478	0.481	-0.6	103	0.00
54 T	cis-1,3-Dichloropropene	0.539	0.525	2.6	100	0.00
55 T	1,1,2-Trichloroethane	0.370	0.357	3.5	103	0.00
56 T	Ethyl methacrylate	0.510	0.564	-10.6	104	0.00
57 T	1,3-Dichloropropane	0.628	0.597	4.9	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.229	0.248	-8.3	103	0.00
59 T	2-Hexanone	0.433	0.431	0.5	98	0.00
60 T	Dibromochloromethane	0.368	0.376	-2.2	105	0.00
61 T	1,2-Dibromoethane	0.381	0.376	1.3	104	0.00
62 S	4-Bromofluorobenzene	0.382	0.422	-10.5	112	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.397	0.355	10.6	99	0.00
65 PM	Chlorobenzene	1.062	1.022	3.8	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.378	0.8	104	0.00
67 C	Ethyl Benzene	1.792	1.804	-0.7#	101	0.00
68 T	m/p-Xylenes	0.683	0.705	-3.2	103	0.00
69 T	o-Xylene	0.666	0.699	-5.0	102	0.00
70 T	Styrene	1.066	1.152	-8.1	102	0.00
71 P	Bromoform	0.314	0.328	-4.5	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.660	3.635	0.7	103	0.00
74 T	N-amyl acetate	1.383	1.426	-3.1	105	0.00
75 P	1,1,2,2-Tetrachloroethane	1.412	1.230	12.9	102	0.00
76 T	1,2,3-Trichloropropane	1.135	1.066	6.1	116	0.00
77 T	Bromobenzene	0.953	0.925	2.9	106	0.00
78 T	n-propylbenzene	4.118	4.136	-0.4	101	0.00
79 T	2-Chlorotoluene	2.570	2.454	4.5	102	0.00
80 T	1,3,5-Trimethylbenzene	3.012	3.008	0.1	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.386	0.352	8.8	95	0.00
82 T	4-Chlorotoluene	2.436	2.352	3.4	101	0.00
83 T	tert-Butylbenzene	2.668	2.700	-1.2	101	0.00
84 T	1,2,4-Trimethylbenzene	2.962	3.033	-2.4	102	0.00
85 T	sec-Butylbenzene	3.687	3.764	-2.1	101	0.00
86 T	p-Isopropyltoluene	2.907	3.180	-9.4	103	0.00
87 T	1,3-Dichlorobenzene	1.740	1.638	5.9	103	0.00
88 T	1,4-Dichlorobenzene	1.741	1.628	6.5	103	0.00
89 T	n-Butylbenzene	2.365	2.482	-4.9	101	0.00
90 T	Hexachloroethane	0.577	0.546	5.4	101	0.00
91 T	1,2-Dichlorobenzene	1.763	1.680	4.7	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.284	0.275	3.2	105	0.00
93 T	1,2,4-Trichlorobenzene	0.876	0.975	-11.3	108	0.00
94 T	Hexachlorobutadiene	0.476	0.474	0.4	106	0.00
95 T	Naphthalene	2.841	5.174	-82.1#	167#	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080122\
Data File : VN073674.D
Acq On : 01 Aug 2022 21:02
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 02 05:07:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.921	0.987	-7.2	106	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6