

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080123\
 Data File : VN078724.D
 Acq On : 01 Aug 2023 18:37
 Operator : JC\MD
 Sample : VN0801WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0801WBL01

Quant Time: Aug 02 04:27:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080123W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 02 04:20:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	98908	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	524307	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	423670	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	215438	29.058	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	96.867%
60) 4-Bromofluorobenzene	12.853	95	139819	22.021	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	73.400%
63) Toluene-d8	10.571	98	637285	30.450	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080123\
Data File : VN078724.D
Acq On : 01 Aug 2023 18:37
Operator : JC\MD
Sample : VN0801WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0801WBL01

Quant Time: Aug 02 04:27:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080123W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 02 04:20:43 2023
Response via : Initial Calibration

