

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080218\
 Data File : VN050292.D
 Acq On : 3 Aug 2018 3:48
 Operator : MD\SY
 Sample : J4304-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EW-A07

Quant Time: Aug 03 04:57:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	600085	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1020117	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	852806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	351517	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	438215	52.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.68%	
35) Dibromofluoromethane	7.59	113	392130	47.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.16%	
50) Toluene-d8	10.09	98	1371837	44.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.60%	
62) 4-Bromofluorobenzene	12.40	95	407695	40.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.88%	
Target Compounds						
					Qvalue	
16) Acetone	3.83	43	13985	2.797	ug/l	91
18) Methyl Acetate	4.33	43	17365	0.656	ug/l #	80
19) Methyl tert-butyl Ether	5.06	73	7758	0.421	ug/l	95
73) Isopropylbenzene	12.25	105	49208	1.805	ug/l	97
78) n-propylbenzene	12.59	91	29295	0.961	ug/l	99
83) tert-Butylbenzene	12.99	119	17302	0.925	ug/l	84
85) sec-Butylbenzene	13.17	105	43399	1.736	ug/l	87
89) n-Butylbenzene	13.62	91	6417	0.370	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080218\
Data File : VN050292.D
Acq On : 3 Aug 2018 3:48
Operator : MD\SY
Sample : J4304-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EW-A07

Quant Time: Aug 03 04:57:07 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

