

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080219\
 Data File : VN057060.D
 Acq On : 2 Aug 2019 12:45
 Operator : JC/SP
 Sample : K4118-23
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-PZ13311-20190731

Quant Time: Aug 14 10:19:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 01:33:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	563968	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	802825	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	660415	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	181725	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	314486	54.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.96%	
35) Dibromofluoromethane	7.58	113	285446	55.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.08%	
50) Toluene-d8	10.08	98	1057029	53.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.86%	
62) 4-Bromofluorobenzene	12.40	95	278601	43.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.34%	
Target Compounds						
64) Tetrachloroethene	10.63	164	6691	1.176	ug/l	96
96) 1,2,3-Trichlorobenzene	15.29	180	2563	2.818	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080219\
Data File : VN057060.D
Acq On : 2 Aug 2019 12:45
Operator : JC/SP
Sample : K4118-23
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-PZ133I1-20190731

Quant Time: Aug 14 10:19:06 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072719W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 01:33:29 2019
Response via : Initial Calibration

