

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080222\
 Data File : VN073682.D
 Acq On : 02 Aug 2022 13:09
 Operator : JC\MD
 Sample : N3974-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Aug 03 02:04:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	356154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	573283	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	543530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	265708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	209475	48.263	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.520%		
35) Dibromofluoromethane	7.951	113	195332	54.261	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	108.520%		
50) Toluene-d8	10.380	98	643284	45.411	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.820%		
62) 4-Bromofluorobenzene	12.674	95	244368	51.362	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	102.720%		
Target Compounds						
16) Acetone	4.240	43	5024	2.155	ug/l	98
95) Naphthalene	15.433	128	5643	0.374	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080222\
Data File : VN073682.D
Acq On : 02 Aug 2022 13:09
Operator : JC\MD
Sample : N3974-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Aug 03 02:04:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

