

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080222\
 Data File : VN073692.D
 Acq On : 02 Aug 2022 17:15
 Operator : JC\MD
 Sample : N3971-18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-26R

Quant Time: Aug 03 02:08:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	362255	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	585864	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	564248	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	269556	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	214080	48.493	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.980%
35) Dibromofluoromethane	7.957	113	200621	54.534	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.060%
50) Toluene-d8	10.380	98	669836	46.270	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.540%
62) 4-Bromofluorobenzene	12.674	95	248119	51.085	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.160%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.304	59	4376	4.596	ug/l #	72
16) Acetone	4.228	43	40711	17.169	ug/l	91
37) Ethyl Acetate	7.339	43	118877	17.216	ug/l	98
40) Benzene	8.386	78	16444	0.966	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080222\
Data File : VN073692.D
Acq On : 02 Aug 2022 17:15
Operator : JC\MD
Sample : N3971-18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-26R

Quant Time: Aug 03 02:08:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

