

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080222\
 Data File : VN073701.D
 Acq On : 02 Aug 2022 20:56
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 03 02:40:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.485	0.463	4.5	95	0.00
3 P	Chloromethane	0.699	0.593	15.2	90	0.00
4 C	Vinyl Chloride	0.848	0.761	10.3#	89	0.00
5 T	Bromomethane	0.492	0.403	18.1	82	0.00
6 T	Chloroethane	0.592	0.516	12.8	94	0.00
7 T	Trichlorofluoromethane	0.874	1.067	-22.1	117	0.00
8 T	Diethyl Ether	0.328	0.336	-2.4	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.484	0.467	3.5	98	0.00
10 T	Methyl Iodide	0.492	0.421	14.4	74	0.00
11 T	Tert butyl alcohol	0.131	0.146	-11.5	104	0.00
12 CM	1,1-Dichloroethene	0.474	0.465	1.9#	99	0.00
13 T	Acrolein	0.130	0.167	-28.5#	114	0.00
14 T	Allyl chloride	0.758	0.729	3.8	93	0.00
15 T	Acrylonitrile	0.364	0.389	-6.9	101	0.00
16 T	Acetone	0.327	0.319	2.4	102	0.00
17 T	Carbon Disulfide	1.258	1.187	5.6	94	0.00
18 T	Methyl Acetate	0.814	0.852	-4.7	105	0.00
19 T	Methyl tert-butyl Ether	1.621	1.771	-9.3	103	0.00
20 T	Methylene Chloride	0.691	0.593	14.2	102	0.00
21 T	trans-1,2-Dichloroethene	0.514	0.516	-0.4	98	0.00
22 T	Diisopropyl ether	1.636	1.701	-4.0	99	0.00
23 T	Vinyl Acetate	1.372	1.496	-9.0	99	0.00
24 P	1,1-Dichloroethane	1.007	0.990	1.7	100	0.00
25 T	2-Butanone	0.508	0.535	-5.3	101	0.00
26 T	2,2-Dichloropropane	0.774	0.508	34.4#	66	0.00
27 T	cis-1,2-Dichloroethene	0.633	0.653	-3.2	101	0.00
28 T	Bromochloromethane	0.423	0.428	-1.2	96	0.00
29 T	Tetrahydrofuran	0.321	0.342	-6.5	98	0.00
30 C	Chloroform	1.049	1.062	-1.2#	101	0.00
31 T	Cyclohexane	0.907	0.843	7.1	94	0.00
32 T	1,1,1-Trichloroethane	0.871	0.888	-2.0	99	0.00
33 S	1,2-Dichloroethane-d4	0.609	0.688	-13.0	112	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.314	0.354	-12.7	113	0.00
36 T	1,1-Dichloropropene	0.447	0.446	0.2	97	0.00
37 T	Ethyl Acetate	0.589	0.588	0.2	96	0.00
38 T	Carbon Tetrachloride	0.454	0.460	-1.3	101	0.00
39 T	Methylcyclohexane	0.535	0.536	-0.2	93	0.00
40 TM	Benzene	1.453	1.453	0.0	99	0.00
41 T	Methacrylonitrile	0.267	0.242	9.4	84	0.00
42 TM	1,2-Dichloroethane	0.490	0.487	0.6	100	0.00
43 T	Isopropyl Acetate	0.796	0.851	-6.9	100	0.00
44 TM	Trichloroethene	0.368	0.370	-0.5	99	0.00
45 C	1,2-Dichloropropane	0.363	0.361	0.6#	99	0.00
46 T	Dibromomethane	0.252	0.259	-2.8	100	0.00
47 T	Bromodichloromethane	0.484	0.497	-2.7	102	0.00
48 T	Methyl methacrylate	0.348	0.375	-7.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080222\
 Data File : VN073701.D
 Acq On : 02 Aug 2022 20:56
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 03 02:40:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.011	-10.0	101	0.00
50 S	Toluene-d8	1.236	1.396	-12.9	110	0.00
51 T	4-Methyl-2-Pentanone	0.571	0.617	-8.1	99	0.00
52 CM	Toluene	0.888	0.934	-5.2#	98	0.00
53 T	t-1,3-Dichloropropene	0.478	0.495	-3.6	95	0.00
54 T	cis-1,3-Dichloropropene	0.539	0.543	-0.7	93	0.00
55 T	1,1,2-Trichloroethane	0.370	0.388	-4.9	101	0.00
56 T	Ethyl methacrylate	0.510	0.608	-19.2	101	0.00
57 T	1,3-Dichloropropane	0.628	0.644	-2.5	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.229	0.259	-13.1	98	0.00
59 T	2-Hexanone	0.433	0.480	-10.9	99	0.00
60 T	Dibromochloromethane	0.368	0.404	-9.8	102	0.00
61 T	1,2-Dibromoethane	0.381	0.401	-5.2	99	0.00
62 S	4-Bromofluorobenzene	0.382	0.464	-21.5	111	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.397	0.375	5.5	95	0.00
65 PM	Chlorobenzene	1.062	1.082	-1.9	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.406	-6.6	102	0.00
67 C	Ethyl Benzene	1.792	1.913	-6.8#	98	0.00
68 T	m/p-Xylenes	0.683	0.743	-8.8	98	0.00
69 T	o-Xylene	0.666	0.738	-10.8	98	0.00
70 T	Styrene	1.066	1.234	-15.8	99	0.00
71 P	Bromoform	0.314	0.356	-13.4	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.660	3.917	-7.0	98	0.00
74 T	N-amyl acetate	1.383	1.527	-10.4	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.412	1.376	2.5	101	0.00
76 T	1,2,3-Trichloropropane	1.135	1.040	8.4	101	0.00
77 T	Bromobenzene	0.953	0.991	-4.0	100	0.00
78 T	n-propylbenzene	4.118	4.401	-6.9	96	0.00
79 T	2-Chlorotoluene	2.570	2.663	-3.6	98	0.00
80 T	1,3,5-Trimethylbenzene	3.012	3.268	-8.5	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.386	0.363	6.0	87	0.00
82 T	4-Chlorotoluene	2.436	2.542	-4.4	97	0.00
83 T	tert-Butylbenzene	2.668	2.957	-10.8	98	0.00
84 T	1,2,4-Trimethylbenzene	2.962	3.262	-10.1	98	0.00
85 T	sec-Butylbenzene	3.687	4.040	-9.6	96	0.00
86 T	p-Isopropyltoluene	2.907	3.334	-14.7	96	0.00
87 T	1,3-Dichlorobenzene	1.740	1.776	-2.1	99	0.00
88 T	1,4-Dichlorobenzene	1.741	1.741	0.0	98	0.00
89 T	n-Butylbenzene	2.365	2.590	-9.5	94	0.00
90 T	Hexachloroethane	0.577	0.593	-2.8	98	0.00
91 T	1,2-Dichlorobenzene	1.763	1.812	-2.8	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.284	0.307	-8.1	104	0.00
93 T	1,2,4-Trichlorobenzene	0.876	1.012	-15.5	99	0.00
94 T	Hexachlorobutadiene	0.476	0.475	0.2	94	0.00
95 T	Naphthalene	2.841	3.567	-25.6#	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080222\
Data File : VN073701.D
Acq On : 02 Aug 2022 20:56
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 03 02:40:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.921	1.063	-15.4	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6