

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080318\
 Data File : VN050337.D
 Acq On : 4 Aug 2018 2:10
 Operator : MD\SY
 Sample : PB111719ZHE#17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111719ZHE#17

Quant Time: Aug 04 05:43:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	627009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1065252	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	928169	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	378524	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	452300	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
35) Dibromofluoromethane	7.59	113	396486	45.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.18%	
50) Toluene-d8	10.09	98	1449295	45.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.64%	
62) 4-Bromofluorobenzene	12.40	95	425923	40.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.90%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	37235	13.97	ug/l	95
20) Methylene Chloride	4.55	84	33899	3.20	ug/l	95
43) Isopropyl Acetate	8.17	43	309109	18.89	ug/l	93
59) 2-Hexanone	10.77	43	21049	8.58	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080318\
Data File : VN050337.D
Acq On : 4 Aug 2018 2:10
Operator : MD\SY
Sample : PB111719ZHE#17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111719ZHE#17

Quant Time: Aug 04 05:43:40 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

