

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080318\
 Data File : VN050342.D
 Acq On : 4 Aug 2018 4:14
 Operator : MD\SY
 Sample : J4313-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-5

Quant Time: Aug 04 06:33:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	690129	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1142853	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	977176	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	380931	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	481048	49.96	ug/l	0.00
Spiked Amount			Recovery	=	99.92%	
35) Dibromofluoromethane	7.59	113	430186	46.10	ug/l	0.00
Spiked Amount			Recovery	=	92.20%	
50) Toluene-d8	10.09	98	1561908	45.53	ug/l	0.00
Spiked Amount			Recovery	=	91.06%	
62) 4-Bromofluorobenzene	12.40	95	441946	39.12	ug/l	0.00
Spiked Amount			Recovery	=	78.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	38738	12.96	ug/l	98
20) Methylene Chloride	4.55	84	416946	48.81	ug/l	99
25) 2-Butanone	6.85	43	12717	3.46	ug/l #	86
30) Chloroform	7.37	83	16737	1.03	ug/l	100
43) Isopropyl Acetate	8.17	43	300748	17.03	ug/l	93
59) 2-Hexanone	10.76	43	17135	7.76	ug/l	66
95) Naphthalene	15.13	128	63475	7.74	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080318\
Data File : VN050342.D
Acq On : 4 Aug 2018 4:14
Operator : MD\SY
Sample : J4313-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-5

Quant Time: Aug 04 06:33:10 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

