

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080318\
 Data File : VN050348.D
 Acq On : 4 Aug 2018 6:43
 Operator : MD\SY
 Sample : J4313-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-11

Quant Time: Aug 06 00:56:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	677550	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1099244	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	951334	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	396861	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	469378	49.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.30%	
35) Dibromofluoromethane	7.59	113	412853	46.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.00%	
50) Toluene-d8	10.09	98	1522969	46.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.32%	
62) 4-Bromofluorobenzene	12.40	95	447728	41.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.42%	
Target Compounds						
16) Acetone	3.82	43	39397	13.59	ug/l	95
20) Methylene Chloride	4.55	84	73508	7.71	ug/l	99
30) Chloroform	7.38	83	19481	1.22	ug/l	99
43) Isopropyl Acetate	8.17	43	219412	12.63	ug/l #	91
59) 2-Hexanone	10.77	43	9247	6.63	ug/l #	58
95) Naphthalene	15.14	128	93684	9.13	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080318\
Data File : VN050348.D
Acq On : 4 Aug 2018 6:43
Operator : MD\SY
Sample : J4313-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-11

Quant Time: Aug 06 00:56:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

