

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080318\
 Data File : VN050349.D
 Acq On : 4 Aug 2018 7:08
 Operator : MD\SY
 Sample : J4313-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-12

Quant Time: Aug 06 00:59:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	655116	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1076551	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	942566	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	407799	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	455604	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
35) Dibromofluoromethane	7.59	113	409025	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
50) Toluene-d8	10.09	98	1478178	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
62) 4-Bromofluorobenzene	12.40	95	446399	41.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.90%	
Target Compounds						
8) Diethyl Ether	3.41	74	6729	1.45	ug/l	98
16) Acetone	3.82	43	44225	16.49	ug/l	99
20) Methylene Chloride	4.55	84	274993	33.52	ug/l	100
43) Isopropyl Acetate	8.17	43	182616	10.56	ug/l #	93
59) 2-Hexanone	10.77	43	6310	6.19	ug/l	64
95) Naphthalene	15.14	128	92402	8.94	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080318\
Data File : VN050349.D
Acq On : 4 Aug 2018 7:08
Operator : MD\SY
Sample : J4313-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-12

Quant Time: Aug 06 00:59:03 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

