

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080318\
 Data File : VN050353.D
 Acq On : 4 Aug 2018 8:47
 Operator : MD\SY
 Sample : J4316-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T40-CLAY-1

Quant Time: Aug 06 01:08:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 26 18:10:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	640003	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1074913	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	927638	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	380881	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	458858	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
35) Dibromofluoromethane	7.59	113	400170	45.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.20%	
50) Toluene-d8	10.09	98	1472319	45.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.26%	
62) 4-Bromofluorobenzene	12.40	95	430314	40.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.00%	
Target Compounds						
16) Acetone	3.82	43	32730	11.42	ug/l	98
20) Methylene Chloride	4.55	84	155130	18.81	ug/l	96
30) Chloroform	7.37	83	16114	1.06	ug/l	97
43) Isopropyl Acetate	8.17	43	290362	17.51	ug/l	93
59) 2-Hexanone	10.77	43	19502	8.31	ug/l #	57
95) Naphthalene	15.14	128	17674	5.34	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080318\
Data File : VN050353.D
Acq On : 4 Aug 2018 8:47
Operator : MD\SY
Sample : J4316-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T40-CLAY-1

Quant Time: Aug 06 01:08:19 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 26 18:10:10 2018
Response via : Initial Calibration

