


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080318\
Data File : VN050303.D
Acq On : 3 Aug 2018 9:52
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072418W.M
Title : SW846 8260
Last Update : Thu Jul 26 18:10:10 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3348

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	13134	PASS
75	95	30	60	50.2	35224	PASS
95	95	100	100	100.0	70133	PASS
96	95	5	9	6.9	4814	PASS
173	174	0.00	2	1.4	880	PASS
174	95	50	100	91.7	64330	PASS
175	174	5	9	6.9	4465	PASS
176	174	95	101	96.3	61933	PASS
177	176	5	9	6.7	4147	PASS