

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080320\
 Data File : VN062731.D
 Acq On : 3 Aug 2020 13:25
 Operator : JC/MD
 Sample : L3539-02 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 I-LDC-5

Quant Time: Aug 04 03:29:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	125474	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	219923	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	220854	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	96864	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	108650	55.74	ug/l	0.00
Spiked Amount			Recovery	=	111.48%	
35) Dibromofluoromethane	7.55	113	74199	51.36	ug/l	0.00
Spiked Amount			Recovery	=	102.72%	
50) Toluene-d8	10.06	98	286943	49.78	ug/l	0.00
Spiked Amount			Recovery	=	99.56%	
62) 4-Bromofluorobenzene	12.38	95	116908	55.47	ug/l	0.00
Spiked Amount			Recovery	=	110.94%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.49	94	175	0.225	ug/l	85
11) Tert butyl alcohol	4.68	59	827	2.667	ug/l #	73
16) Acetone	3.78	43	20931	28.536	ug/l	94
17) Carbon Disulfide	4.00	76	232737	52.723	ug/l	99
18) Methyl Acetate	4.28	43	1943	1.046	ug/l #	70
20) Methylene Chloride	4.49	84	847	0.481	ug/l #	37
25) 2-Butanone	6.80	43	3637	3.256	ug/l	98
30) Chloroform	7.33	83	1348	0.456	ug/l	94
31) Cyclohexane	7.62	56	2961	1.031	ug/l #	1
36) 1,1-Dichloropropene	7.63	75	10552	4.501	ug/l #	51
37) Ethyl Acetate	6.80	43	3637	1.439	ug/l #	69
38) Carbon Tetrachloride	7.62	117	13075	5.803	ug/l #	18
40) Benzene	8.00	78	19992	2.817	ug/l	96
52) Toluene	10.12	92	3866	0.938	ug/l	95
55) 1,1,2-Trichloroethane	10.38	97	1517	0.949	ug/l #	13
56) Ethyl methacrylate	10.23	69	781	2.031	ug/l #	1
59) 2-Hexanone	10.72	43	420	3.179	ug/l #	26
67) Ethyl Benzene	11.49	91	1864	0.220	ug/l #	80
68) m/p-Xylenes	11.59	106	3195	1.047	ug/l	92
69) o-Xylene	11.92	106	1278	0.441	ug/l	82
71) Bromoform	12.37	173	263	0.201	ug/l #	29
76) 1,2,3-Trichloropropane	12.38	75	64466	26.147	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.38	75	64466	72.787	ug/l #	12
84) 1,2,4-Trimethylbenzene	13.01	105	2475	0.397	ug/l	96
93) 1,2,4-Trichlorobenzene	14.89	180	373	3.319	ug/l #	84
95) Naphthalene	15.10	128	258831	59.328	ug/l	99
96) 1,2,3-Trichlorobenzene	15.29	180	496	0.341	ug/l #	59


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080320\
Data File : VN062731.D
Acq On : 3 Aug 2020 13:25
Operator : JC/MD
Sample : L3539-02 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
I-LDC-5

Quant Time: Aug 04 03:29:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. -0.00 min

Lab File: VN062731.D

Acq: 3 Aug 2020 13:25

Instrument :

MSVOA_N

ClientSampled :


I-LDC-5

Tot Ion:168 Resp: 125474

Ion Ratio Lower Upper

168 100

99 65.7 48.4 72.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.63

60000

50000

40000



30000

20000

10000

0

Time--> 7.50 7.60 7.70



#5

Bromomethane

Concen: 0.225 ug/l

RT: 2.49 min Scan# 223

Delta R.T. -0.03 min

Lab File: VN062731.D


Acq: 3 Aug 2020 13:25

Tgt Ion: 94 Resp: 175

Ion Ratio Lower Upper

94 100

96 109.9 76.6 114.8



Abundance Ion 93.90 (93.60 to 94.60): VNO

Ion 95.90 (95.60 to 96.60): VNO

2.49

400


300


200

100

0

Time--> 2.47 2.48 2.49 2.50









#11

Tert butyl alcohol
 Concen: 2.667 ug/l
 RT: 4.68 min Scan# 905
 Delta R.T. -0.05 min
 Lab File: VN062731.D
 Acq: 3 Aug 2020 13:25

Instrument :
 MSVOA_N
 ClientSampleId :
 I-LDC-5




Tot Ion: 59 Resp: 827
 Ion Ratio Lower Upper
 59 100
 57 0.0 8.2 12.2#




#16

Acetone
 Concen: 28.536 ug/l
 RT: 3.78 min Scan# 624
 Delta R.T. 0.01 min
 Lab File: VN062731.D
 Acq: 3 Aug 2020 13:25

Tgt Ion: 43 Resp: 20931
 Ion Ratio Lower Upper
 43 100
 58 27.6 24.6 37.0









#17
Carbon Disulfide
Concen: 52.723 ug/l
RT: 4.00 min Scan# 693
Delta R.T. -0.01 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25




Instrument :
MSVOA_N
ClientSampleId :
I-LDC-5


Tot Ion: 76 Resp: 232737
Ion Ratio Lower Upper
76 100
78 8.9 7.3 10.9



#18
Methyl Acetate
Concen: 1.046 ug/l
RT: 4.28 min Scan# 780
Delta R.T. -0.00 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25

Tgt Ion: 43 Resp: 1943
Ion Ratio Lower Upper
43 100
74 6.6 16.6 24.8#









#20
Methvlene Chloride
Concen: 0.481 ug/l
RT: 4.49 min Scan# 846
Delta R.T. -0.01 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25




Instrument :
MSVOA_N
ClientSampled :
I-LDC-5


Tot Ion: 84 Resp: 847
Ion Ratio Lower Upper
84 100
49 38.2 109.4 164.0#
51 0.0 33.6 50.4#
86 48.1 51.4 77.2#



#25
2-Butanone
Concen: 3.256 ug/l
RT: 6.80 min Scan# 1563
Delta R.T. 0.01 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25

Tgt Ion: 43 Resp: 3637
Ion Ratio Lower Upper
43 100
72 25.0 19.0 28.6









#30
Chloroform
Concen: 0.456 ug/l
RT: 7.33 min Scan# 1727
Delta R.T. -0.00 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25




Instrument :
MSVOA_N
ClientSampleId :
I-LDC-5


Tot Ion: 83 Resp: 1348
Ion Ratio Lower Upper
83 100
85 69.8 52.0 78.0



#31
Cyclohexane
Concen: 1.031 ug/l
RT: 7.62 min Scan# 1819
Delta R.T. 0.01 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25

Tgt Ion: 56 Resp: 2961
Ion Ratio Lower Upper
56 100
69 172.8 25.3 37.9#
84 116.5 61.0 91.4#





#33

1,2-Dichloroethane-d4

Concen: 55.741 ug/l

RT: 7.99 min Scan# 1932

Delta R.T. -0.00 min

Lab File: VN062731.D

Acq: 3 Aug 2020 13:25

Instrument :

MSVOA_N

ClientSampleId :


I-LDC-5

Tot Ion: 65 Resp: 108650

Ion Ratio Lower Upper




65 100

67 49.2 0.0 106.0



Abundance Ion 64.90 (64.60 to 65.60): VN062731.D (-1840) (-)

Ion 66.90 (66.60 to 67.60): VN062731.D (-1840) (-)



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2108

Delta R.T. -0.00 min

Lab File: VN062731.D

Acq: 3 Aug 2020 13:25


Tgt Ion: 114 Resp: 219923

Ion Ratio Lower Upper

114 100

63 23.5 0.0 48.4



88 17.9 0.0 33.8




Abundance Ion 113.90 (113.60 to 114.60): VN062731.D (-2015) (-)

Ion 63.00 (62.70 to 63.70): VN062731.D (-2015) (-)

Ion 87.90 (87.60 to 88.60): VN062731.D (-2015) (-)









#35
 Dibromofluoromethane
 Concen: 51.364 ug/l
 RT: 7.55 min Scan# 1796
 Delta R.T. -0.00 min
 Lab File: VN062731.D
 Acq: 3 Aug 2020 13:25

Instrument :
 MSVOA_N
 ClientSampleId :
 I-LDC-5

Tot Ion	Ratio	Lower	Upper
113	100		
111	101.0	82.7	124.1
192	17.5	14.8	22.2




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#36
 1,1-Dichloropropene
 Concen: 4.501 ug/l
 RT: 7.63 min Scan# 1820
 Delta R.T. -0.13 min
 Lab File: VN062731.D
 Acq: 3 Aug 2020 13:25

Tgt Ion	Ratio	Lower	Upper
75	100		
110	8.4	16.0	48.0#
77	0.0	24.6	36.8#



Abundance Ion 74.90 (74.60 to 75.60): VNO
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): VNO









#37
Ethyl Acetate
Concen: 1.439 ug/l
RT: 6.80 min Scan# 1563
Delta R.T. -0.08 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25

Instrument :
MSVOA_N
ClientSampleId :
I-LDC-5

Tot Ion: 43 Resp: 3637
Ion Ratio Lower Upper
43 100
61 0.0 11.3 16.9#
70 0.0 7.7 11.5#




Abundance Ion 43.00 (42.70 to 43.70): VN0
Ion 60.90 (60.60 to 61.60): VN0
Ion 70.00 (69.70 to 70.70): VN0






#38
Carbon Tetrachloride
Concen: 5.803 ug/l
RT: 7.62 min Scan# 1819
Delta R.T. -0.11 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25

Tgt Ion: 117 Resp: 13075
Ion Ratio Lower Upper
117 100
119 5.8 74.8 112.2#
121 0.0 23.0 34.6#



Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V









#40
Benzene
Concen: 2.817 ug/l
RT: 8.00 min Scan# 1937
Delta R.T. -0.00 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25




Instrument :
MSVOA_N
ClientSampleId :
I-LDC-5


Tot Ion: 78 Resp: 19992
Ion Ratio Lower Upper
78 100
77 25.5 18.7 28.1



#50
Toluene-d8
Concen: 49.784 ug/l
RT: 10.06 min Scan# 2577
Delta R.T. -0.00 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25

Tgt Ion: 98 Resp: 286943
Ion Ratio Lower Upper
98 100
100 62.7 50.9 76.3





#52

Toluene

Concen: 0.938 ug/l

RT: 10.12 min Scan# 2597

Delta R.T. -0.00 min

Lab File: VN062731.D

Acq: 3 Aug 2020 13:25

Instrument :

MSVOA_N

ClientSampleId :





I-LDC-5

Tot Ion: 92 Resp: 3866

Ion Ratio Lower Upper

92 100

91 181.4 139.4 209.2



#55

1,1,2-Trichloroethane

Concen: 0.949 ug/l

RT: 10.38 min Scan# 2677

Delta R.T. -0.15 min

Lab File: VN062731.D

Acq: 3 Aug 2020 13:25

Tgt Ion: 97 Resp: 1517




Ion Ratio Lower Upper


97 100

83 0.0 73.7 110.5#

85 0.0 48.5 72.7#

99 0.0 50.8 76.2#





#56

Ethyl methacrylate

Concen: 2.031 ug/l

RT: 10.23 min Scan# 2629

Delta R.T. -0.17 min

Lab File: VN062731.D

Acq: 3 Aug 2020 13:25

Instrument :

MSVOA_N

ClientSampled :

I-LDC-5





Tot Ion: 69 Resp: 781

Ion Ratio Lower Upper

69 100

41 0.0 57.2 85.8#

39 141.6 28.6 43.0#



#59

2-Hexanone

Concen: 3.179 ug/l

RT: 10.72 min Scan# 2781

Delta R.T. -0.01 min

Lab File: VN062731.D




Acq: 3 Aug 2020 13:25


Tgt Ion: 43 Resp: 420

Ion Ratio Lower Upper

43 100

58 0.0 26.0 78.0#





#62

4-Bromofluorobenzene

Concen: 55.473 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.00 min

Lab File: VN062731.D

Acq: 3 Aug 2020 13:25

Instrument :

MSVOA_N

ClientSampled :

I-LDC-5


Tot Ion: 95 Resp: 116908

Ion Ratio Lower Upper

95 100

174 72.7 0.0 144.6




176 70.0 0.0 139.4



Abundance Ion 94.90 (94.60 to 95.60): VN062731.D (-3205) (-)

Ion 173.80 (173.50 to 174.50): VN062731.D (-3205) (-)

Ion 175.80 (175.50 to 176.50): VN062731.D (-3205) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.38 min Scan# 2988

Delta R.T. -0.00 min

Lab File: VN062731.D

Acq: 3 Aug 2020 13:25


Tgt Ion: 117 Resp: 220854

Ion Ratio Lower Upper

117 100

82 62.6 49.0 73.6



119 32.1 25.4 38.2




Abundance Ion 116.90 (116.60 to 117.60): VN062731.D (-2895) (-)

Ion 82.00 (81.70 to 82.70): VN062731.D (-2895) (-)

Ion 118.90 (118.60 to 119.60): VN062731.D (-2895) (-)









#67
Ethyl Benzene
Concen: 0.220 ug/l
RT: 11.49 min Scan# 3021
Delta R.T. 0.00 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25




Instrument :
MSVOA_N
ClientSampled :
I-LDC-5


Tot Ion: 91 Resp: 1864
Ion Ratio Lower Upper
91 100
106 19.1 23.8 35.8#



#68
m/p-Xylenes
Concen: 1.047 ug/l
RT: 11.59 min Scan# 3053
Delta R.T. -0.01 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25

Tgt Ion: 106 Resp: 3195
Ion Ratio Lower Upper
106 100
91 227.3 171.4 257.2









#69
o-Xylene
Concen: 0.441 ug/l
RT: 11.92 min Scan# 3157
Delta R.T. 0.00 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25




Instrument :
MSVOA_N
ClientSampled :
I-LDC-5


Tot Ion:106 Resp: 1278
Ion Ratio Lower Upper
106 100
91 257.0 113.8 341.2



#71
Bromoform
Concen: 0.201 ug/l
RT: 12.37 min Scan# 3295
Delta R.T. 0.27 min
Lab File: VN062731.D
Acq: 3 Aug 2020 13:25

Tgt Ion:173 Resp: 263
Ion Ratio Lower Upper
173 100
175 0.0 24.1 72.2#
254 0.0 0.0 0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3591

Delta R.T. 0.00 min

Lab File: VN062731.D

Acq: 3 Aug 2020 13:25

Instrument :

MSVOA_N

ClientSampleId :

I-LDC-5


Tot Ion:152 Resp: 96864

Ion Ratio Lower Upper

152 100

115 64.0 31.8 95.4




150 157.3 0.0 351.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 26.147 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.15 min

Lab File: VN062731.D


Acq: 3 Aug 2020 13:25

Tgt Ion: 75 Resp: 64466

Ion Ratio Lower Upper



75 100


77 1.1 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): V

Ion 77.00 (76.70 to 77.70): V









#81
 trans-1,4-Dichloro-2-butene
 Concen: 72.787 ug/l
 RT: 12.38 min Scan# 3298
 Delta R.T. 0.10 min
 Lab File: VN062731.D
 Acq: 3 Aug 2020 13:25




Instrument :
 MSVOA_N
 ClientSampled :
 I-LDC-5


Tot Ion: 75 Resp: 64466
 Ion Ratio Lower Upper
 75 100
 53 0.0 75.4 113.0#
 89 0.0 34.5 51.7#



#84
 1,2,4-Trimethylbenzene
 Concen: 0.397 ug/l
 RT: 13.01 min Scan# 3496
 Delta R.T. -0.00 min
 Lab File: VN062731.D
 Acq: 3 Aug 2020 13:25

Tgt Ion: 105 Resp: 2475
 Ion Ratio Lower Upper
 105 100
 120 41.4 22.1 66.3









#93
 1,2,4-Trichlorobenzene
 Concen: 3.319 ug/l
 RT: 14.89 min Scan# 4080
 Delta R.T. 0.01 min
 Lab File: VN062731.D
 Acq: 3 Aug 2020 13:25




Instrument :
 MSVOA_N
 ClientSampleId :
 I-LDC-5


Tot Ion:180 Resp: 373
 Ion Ratio Lower Upper
 180 100
 182 92.5 47.8 143.3
 145 0.0 15.3 45.9#



#95
 Naphthalene
 Concen: 59.328 ug/l
 RT: 15.10 min Scan# 4146
 Delta R.T. -0.00 min
 Lab File: VN062731.D
 Acq: 3 Aug 2020 13:25

Tgt Ion:128 Resp: 258831
 Ion Ratio Lower Upper
 128 100
 127 12.9 10.4 15.6
 129 10.1 8.7 13.1






#96
 1,2,3-Trichlorobenzene
 Concen: 0.341 ug/l
 RT: 15.29 min Scan# 4203
 Delta R.T. 0.00 min
 Lab File: VN062731.D
 Acq: 3 Aug 2020 13:25

Instrument :
 MSVOA_N
 ClientSampleId :
 I-LDC-5

Tot Ion	Ratio	Lower	Upper
180	100		
182	62.5	48.1	144.3
145	0.0	16.7	50.1



Abundance Ion 179.80 (179.50 to 180.50): V
 Ion 181.80 (181.50 to 182.50): V
 Ion 144.90 (144.60 to 145.60): V

