

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080320\
 Data File : VN062727.D
 Acq On : 3 Aug 2020 11:33
 Operator : JC/MD
 Sample : VN0803WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0803WBL01

Quant Time: Aug 04 07:36:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	141206	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	244605	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	238180	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	88188	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	120538	54.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.90%	
35) Dibromofluoromethane	7.55	113	82800	51.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.06%	
50) Toluene-d8	10.06	98	315283	49.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.36%	
62) 4-Bromofluorobenzene	12.38	95	115665	49.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080320\
Data File : VN062727.D
Acq On : 3 Aug 2020 11:33
Operator : JC/MD
Sample : VN0803WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0803WBL01

Quant Time: Aug 04 07:36:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

