

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080320\
 Data File : VN062732.D
 Acq On : 3 Aug 2020 13:49
 Operator : JC/MD
 Sample : L3539-03DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 I-LGC-5DL

Quant Time: Aug 04 07:56:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	126523	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	221102	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	224778	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	89302	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	107368	54.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.26%	
35) Dibromofluoromethane	7.55	113	74231	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
50) Toluene-d8	10.06	98	293479	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
62) 4-Bromofluorobenzene	12.38	95	112155	52.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.86%	
Target Compounds						
16) Acetone	3.78	43	18996	25.683	ug/l	95
17) Carbon Disulfide	4.00	76	116603	26.196	ug/l	96
40) Benzene	8.00	78	1894	0.265	ug/l #	84
95) Naphthalene	15.11	128	16102	7.511	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080320\
Data File : VN062732.D
Acq On : 3 Aug 2020 13:49
Operator : JC/MD
Sample : L3539-03DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
I-LGC-5DL

Quant Time: Aug 04 07:56:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

