

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080320\
 Data File : VN062742.D
 Acq On : 3 Aug 2020 17:50
 Operator : JC/MD
 Sample : L3537-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JTY1-WC-S-BL-IB

Quant Time: Aug 04 05:16:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	119144	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	213790	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	214939	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	78105	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	102951	55.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.24%	
35) Dibromofluoromethane	7.55	113	69877	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
50) Toluene-d8	10.06	98	280715	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
62) 4-Bromofluorobenzene	12.38	95	101797	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
Target Compounds						
16) Acetone	3.79	43	10649	15.290	ug/l	95
18) Methyl Acetate	4.28	43	4348	2.466	ug/l #	73
20) Methylene Chloride	4.50	84	7071	4.233	ug/l	94
43) Isopropyl Acetate	8.13	43	33817	8.540	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080320\
Data File : VN062742.D
Acq On : 3 Aug 2020 17:50
Operator : JC/MD
Sample : L3537-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JTY1-WC-S-BL-IB

Quant Time: Aug 04 05:16:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

