

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080320\
 Data File : VN062744.D
 Acq On : 3 Aug 2020 18:38
 Operator : JC/MD
 Sample : L3537-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JTY1-WC-S-01

Quant Time: Aug 04 05:20:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 18:48:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	119260	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	220559	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	222858	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	79438	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	105491	56.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.88%	
35) Dibromofluoromethane	7.55	113	73300	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
50) Toluene-d8	10.06	98	287352	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
62) 4-Bromofluorobenzene	12.38	95	105864	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
Target Compounds						
16) Acetone	3.78	43	9017	12.934	ug/l	95
18) Methyl Acetate	4.28	43	5001	2.834	ug/l #	85
20) Methylene Chloride	4.50	84	6613	3.955	ug/l	96
43) Isopropyl Acetate	8.13	43	38630	9.456	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080320\
Data File : VN062744.D
Acq On : 3 Aug 2020 18:38
Operator : JC/MD
Sample : L3537-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JTY1-WC-S-01

Quant Time: Aug 04 05:20:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N072120W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 18:48:59 2020
Response via : Initial Calibration

