

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080321\
 Data File : VN067980.D
 Acq On : 3 Aug 2021 14:33
 Operator : JC/MD
 Sample : M3269-01 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Aug 04 03:38:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	334811	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	636878	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	612510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	216741	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	280464	59.172	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	118.340%	
35) Dibromofluoromethane	8.021	113	211693	57.035	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	114.080%	
50) Toluene-d8	10.443	98	881667	51.357	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	102.720%	
62) 4-Bromofluorobenzene	12.734	95	286681	46.149	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	92.300%	
Target Compounds						
						Qvalue
16) Acetone	4.299	43	3146	1.819	ug/l	95
95) Naphthalene	15.507	128	24507	5.663	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080321\
Data File : VN067980.D
Acq On : 3 Aug 2021 14:33
Operator : JC/MD
Sample : M3269-01 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Aug 04 03:38:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

