

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080321\
 Data File : VN067985.D
 Acq On : 3 Aug 2021 16:36
 Operator : JC/MD
 Sample : PB138159ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138159ZHE#01

Quant Time: Aug 04 03:40:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	301921	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	585668	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	562838	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	198655	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	239478	56.029	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.060%
35) Dibromofluoromethane	8.024	113	182715	53.532	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.060%
50) Toluene-d8	10.443	98	744382	47.468	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.940%
62) 4-Bromofluorobenzene	12.734	95	243812	43.054	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.100%
Target Compounds						
20) Methylene Chloride	5.101	84	17047	4.475	ug/l	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080321\
Data File : VN067985.D
Acq On : 3 Aug 2021 16:36
Operator : JC/MD
Sample : PB138159ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138159ZHE#01

Quant Time: Aug 04 03:40:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

