

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080321\
 Data File : VN067989.D
 Acq On : 3 Aug 2021 18:14
 Operator : JC/MD
 Sample : M3269-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-5

Manual Integrations
 APPROVED

MMDadoda
 8/5/2021 6:40:58 PM

Quant Time: Aug 04 03:41:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	278662	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	539959	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	526657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	186643	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	227494	57.667	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 115.340%			
35) Dibromofluoromethane	8.024	113	167089	53.098	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 106.200%			
50) Toluene-d8	10.443	98	692683	47.874	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 95.740%			
62) 4-Bromofluorobenzene	12.734	95	230669	44.051	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 88.100%			
Target Compounds					Qvalue	
16) Acetone	4.304	43	10177m	7.070	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080321\
Data File : VN067989.D
Acq On : 3 Aug 2021 18:14
Operator : JC/MD
Sample : M3269-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-5

Manual Integrations
APPROVED

MMDadoda
8/5/2021 6:40:58 PM

Quant Time: Aug 04 03:41:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

