

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080321\
 Data File : VN067992.D
 Acq On : 3 Aug 2021 19:27
 Operator : JC/MD
 Sample : M3261-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41936

Manual Integrations
 APPROVED

MMDadoda
 8/5/2021 6:41:02 PM

Quant Time: Aug 04 03:42:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	306889	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	582738	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	591848	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	247493	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	236657	54.472	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 108.940%			
35) Dibromofluoromethane	8.024	113	186204	54.829	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 109.660%			
50) Toluene-d8	10.443	98	755974	48.370	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 96.740%			
62) 4-Bromofluorobenzene	12.734	95	271791	47.637	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 95.280%			
Target Compounds						Qvalue
16) Acetone	4.293	43	51271	32.343	ug/l	93
17) Carbon Disulfide	4.524	76	47843	5.669	ug/l	98
25) 2-Butanone	7.346	43	26505	11.525	ug/l	94
42) 1,2-Dichloroethane	8.528	62	2978	0.505	ug/l	87
52) Toluene	10.508	92	4104166	395.811	ug/l	99
67) Ethyl Benzene	11.849	91	20913	0.976	ug/l	91
68) m/p-Xylenes	11.956	106	25634	6.688	ug/l	89
69) o-Xylene	12.286	106	14274	1.811	ug/l	86
78) n-propylbenzene	12.921	91	8638m	0.379	ug/l	
79) 2-Chlorotoluene	13.010	91	1011818	72.858	ug/l	92
82) 4-Chlorotoluene	13.106	91	1048739	75.216	ug/l	93
84) 1,2,4-Trimethylbenzene	13.369	105	85416	5.387	ug/l	99
86) p-Isopropyltoluene	13.616	119	5616	0.371	ug/l	94
89) n-Butylbenzene	13.940	91	3360	0.259	ug/l #	82
95) Naphthalene	15.507	128	23760	5.279	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080321\
Data File : VN067992.D
Acq On : 3 Aug 2021 19:27
Operator : JC/MD
Sample : M3261-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41936

Manual Integrations
APPROVED

MMDadoda
8/5/2021 6:41:02 PM

Quant Time: Aug 04 03:42:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

