

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080322\
 Data File : VN073709.D
 Acq On : 03 Aug 2022 14:11
 Operator : JC\MD
 Sample : N3971-10 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-14

Quant Time: Aug 04 05:33:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	364472	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	584838	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	559039	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	303683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	213992	48.178	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 96.360%			
35) Dibromofluoromethane	7.951	113	195714	53.293	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.580%			
50) Toluene-d8	10.380	98	652751	45.169	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 90.340%			
62) 4-Bromofluorobenzene	12.674	95	264613	53.975	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 107.940%			
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	4897	2.053	ug/l	92
39) Methylcyclohexane	9.392	83	3086	0.493	ug/l #	87
40) Benzene	8.392	78	208584	12.272	ug/l	99
52) Toluene	10.445	92	78089	7.522	ug/l	97
67) Ethyl Benzene	11.786	91	504211	25.169	ug/l	99
68) m/p-Xylenes	11.892	106	914371	119.674	ug/l	100
69) o-Xylene	12.222	106	368280	49.440	ug/l	98
73) Isopropylbenzene	12.521	105	32221	1.449	ug/l	98
78) n-propylbenzene	12.863	91	79351	3.172	ug/l	99
80) 1,3,5-Trimethylbenzene	12.998	105	125214	6.844	ug/l	98
84) 1,2,4-Trimethylbenzene	13.310	105	470530	26.157	ug/l	99
95) Naphthalene	15.439	128	81894	4.746	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080322\
Data File : VN073709.D
Acq On : 03 Aug 2022 14:11
Operator : JC\MD
Sample : N3971-10 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-14

Quant Time: Aug 04 05:33:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

