

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080322\
 Data File : VN073711.D
 Acq On : 03 Aug 2022 15:00
 Operator : JC\MD
 Sample : N3971-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-15

Quant Time: Aug 04 05:34:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	360730	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	569981	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	556331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	260557	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	212162	48.262	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.520%
35) Dibromofluoromethane	7.951	113	194936	54.465	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.920%
50) Toluene-d8	10.381	98	647370	45.964	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.920%
62) 4-Bromofluorobenzene	12.675	95	241832	51.163	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.320%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	58077	24.597	ug/l	99
29) Tetrahydrofuran	7.622	42	10969	4.731	ug/l	96
37) Ethyl Acetate	7.345	43	66149	9.847	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080322\
Data File : VN073711.D
Acq On : 03 Aug 2022 15:00
Operator : JC\MD
Sample : N3971-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-15

Quant Time: Aug 04 05:34:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

