

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080322\
 Data File : VN073713.D
 Acq On : 03 Aug 2022 15:49
 Operator : JC\MD
 Sample : N3972-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPN-1

Quant Time: Aug 04 05:35:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	344998	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	557183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	524779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	240504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	206265	49.060	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.120%
35) Dibromofluoromethane	7.951	113	191246	54.661	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.320%
50) Toluene-d8	10.380	98	625866	45.458	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.920%
62) 4-Bromofluorobenzene	12.674	95	227664	49.575	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.160%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	58611	25.955	ug/l	99
25) 2-Butanone	7.275	43	6740	1.922	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080322\
Data File : VN073713.D
Acq On : 03 Aug 2022 15:49
Operator : JC\MD
Sample : N3972-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPN-1

Quant Time: Aug 04 05:35:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

