

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080322\
 Data File : VN073714.D
 Acq On : 03 Aug 2022 16:14
 Operator : JC\MD
 Sample : N3972-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPN-4

Quant Time: Aug 04 05:35:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	347905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	557824	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	529448	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	237556	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	211343	49.848	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.700%
35) Dibromofluoromethane	7.957	113	194649	55.570	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.140%
50) Toluene-d8	10.380	98	640484	46.466	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.940%
62) 4-Bromofluorobenzene	12.674	95	230143	49.979	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.960%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	56611	24.860	ug/l	97
29) Tetrahydrofuran	7.616	42	8165	3.652	ug/l	93
37) Ethyl Acetate	7.345	43	43776	6.658	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080322\
Data File : VN073714.D
Acq On : 03 Aug 2022 16:14
Operator : JC\MD
Sample : N3972-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPN-4

Quant Time: Aug 04 05:35:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

