

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080322\
 Data File : VN073722.D
 Acq On : 03 Aug 2022 19:31
 Operator : JC\MD
 Sample : N3972-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2022-07-28

Quant Time: Aug 04 05:38:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	331408	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	538832	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	498618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	224042	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	203864	50.477	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.960%	
35) Dibromofluoromethane	7.957	113	188886	55.825	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	111.660%	
50) Toluene-d8	10.380	98	609977	45.813	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	91.620%	
62) 4-Bromofluorobenzene	12.674	95	213561	48.323	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	96.640%	
Target Compounds						
16) Acetone	4.240	43	3704	1.708	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080322\
Data File : VN073722.D
Acq On : 03 Aug 2022 19:31
Operator : JC\MD
Sample : N3972-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2022-07-28

Quant Time: Aug 04 05:38:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

