

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080323\
 Data File : VN078747.D
 Acq On : 03 Aug 2023 12:55
 Operator : JC\MD
 Sample : VN0803WBL03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0803WBL03

Quant Time: Aug 04 06:30:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080123W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 02 04:51:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	149449	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	808015	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.870	117	409821	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	323006	28.833	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	96.100%
60) 4-Bromofluorobenzene	12.859	95	126186	20.546	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	68.500%
63) Toluene-d8	10.570	98	951968	47.022	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	156.733%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080323\
Data File : VN078747.D
Acq On : 03 Aug 2023 12:55
Operator : JC\MD
Sample : VN0803WBL03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0803WBL03

Quant Time: Aug 04 06:30:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080123W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 02 04:51:47 2023
Response via : Initial Calibration

